RETURN TO TECHNICAL SERVICE BULLETIN TABLE OF CONTENTS
RETURN TO CD-ROM TABLE OF CONTENTS

TECHNICAL SERVICE BULLETIN #7

Product: Fabius GS

Date: October 3, 2003

Re: Software Release - 1.39

Reference Doc: Product Bulletin # 153, SP00255, SP00269

Reason/Rational: Enhancement

Solution/Procedure: If applicable, update 1.37 (field) Software kits to 1.39. Perform Software 1.39 upgrade on all Fabius GS units. Complete the Return Form (attached) and return to DMI/DS within 7 days of account completion.

Devices affected: All

Timing/When: At customer request, next service or Periodic Manufacturer's Service

Additional Info: Software 1.39 supersedes 1.37 upon release. Therefore, all 1.37 kits located in the field shall be updated to 1.39 by ordering P/N 4117937-005 - Software 1.39 CD. Or, machine updates from 1.3n to 1.39 shall be completed using Software Kit - 4118364

Note: A new version of Send Image (utility) file is included on the Software CD which must be used to perform the upgrade. Therefore, prior to downloading any software update your local copy of the Send Image file. See applicable Service Procedure's for details.

Cost: Software upgrades to 1.39 are available to Fabius GS customers at no charge.

Ordering Info/Parts: Contact your DMI Customer Service Representative to order all software kits.

Distribution: DrägerService & All Authorized Service Organizations in North America.

DrägerService
3122 Commerce Drive
Telford, PA 18969
Tel: (215) 721-5402
(800) 543-5047
Fax: (215) 723-5935
Web: www.draegermedical.com
DrägerService is a division of
Draeger Medical, Inc.

RETURN TO TECHNICAL SERVICE BULLETIN TABLE OF CONTENTS
RETURN TO CD-ROM TABLE OF CONTENTS

Return Form - Software 1.39 Upgrade

Distributor:	Customer:
Country:	Address:
	City:

Machine S/N	Machine S/N	Machine S/N
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

Fax To:
DrägerService
Attn: Al Myers
(215) 721-5784

Date of Update _____
Signature (TSR) _____
Name (Print) _____

DrägerService
3122 Commerce Drive
Telford, PA 18969
Tel: (215) 721-5402
(800) 543-5047
Fax: (215) 723-5935
Web: www.draegermedical.com
DrägerService is a division of
Draeger Medical, Inc.